

**Corning Incorporated
Life Sciences**

Registered
ISO 9001:2008

Product Description

Catalog Number: 61220-1L

Product Description: PYREXPLUS® Bottle, Aspirator, Protective Coating, Tubulation, 1000 mL
Tooled for Rubber Stopper #6

Component Materials:

Glass - Type I Class A - Low-Expansion Borosilicate Glass,
per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)
Coating - Polyvinyl chloride (PVC)

Attributes:

Capacity - 1000 mL Approximate Below Neck
Product Dimensions - 110 mm OD x 201 mm High
Tubulation - 10 mm OD

Autoclaving:

This product can be autoclaved at 121°C and 15 psi for 15 minutes maximum. Drying cycle time should be less than 15 minutes at no more than 110°C. Steam sterilization can cause cloudiness of the coating which should reduce as trapped moisture evaporates.

Heating / Cooling:

This product is intended for use at room temperature. The coating can be damaged and lose its functional properties if exposed to temperature below -20°C and if continuously exposed to temperature > 80°C. The coating melts at 130°C.
This product cannot be heated by direct contact or flame.

Vacuum / Pressure:

This product is not designed for and is not rated for vacuum or pressure applications.

Sterilization:

Shipped Not Sterile

Lot Number Designation:

Each production lot is uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Revision Date:

12 AUG 2013

Rev No: 1